



Pacesetter™ Air Walker

Single Patient Use Only

Rx Only



IMPORTANT INFORMATION

Please read all instructions before use.

INDICATIONS


Grade 2 and 3 ankle sprains; stable fractures of tibia, fibula, talus and tarsus; Post-operative immobilization (soft tissue reconstructions and open reduction internal fixation); Metatarsal fractures

PRECAUTIONS

- Consult your medical professional immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product.

INSTRUCTIONS FOR USE

- Remove inner liner/wrap from brace.
 - Place foot and lower leg into liner.




NOTE: Remove outer covering from hook strip on inside of liner (on leg and foot portion). Wrap liner snugly around limb. Fasten to outside opposite side of liner.

- Insert wrapped limb into brace frame. Carefully center the uprights over the ankle to ensure a comfortable fit and proper support.




NOTE: Remove plastic covers from uprights and secure hook on uprights to liner.

- Begin fastening the boot straps from the top of foot and working up the leg, using loop lock closures.



- To inflate liner, turn the inflate/deflate dial clockwise. Do not overtighten. Next, pump inflator bulb until desired compression is achieved.
- To deflate liner, turn the inflate/deflate dial counter-clockwise to release air pressure. Turn inflate/deflate dial clockwise when desired compression is achieved.



- Check for adequate circulation. Readjust air amount as needed for a snug and comfortable fit.

WASHING INSTRUCTIONS

Hand wash, lukewarm water, mild detergent, air dry.



Manufactured by DeRoyal
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part#1-1578
©2009 DeRoyal. All Rights Reserved.
Revised 6/09



Pacesetter™ Air Walker

Single Patient Use Only

Rx Only



IMPORTANT INFORMATION

Please read all instructions before use.

INDICATIONS


Grade 2 and 3 ankle sprains; stable fractures of tibia, fibula, talus and tarsus; Post-operative immobilization (soft tissue reconstructions and open reduction internal fixation); Metatarsal fractures

PRECAUTIONS

- Consult your medical professional immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product.

INSTRUCTIONS FOR USE

- Remove inner liner/wrap from brace.
 - Place foot and lower leg into liner.




NOTE: Remove outer covering from hook strip on inside of liner (on leg and foot portion). Wrap liner snugly around limb. Fasten to outside opposite side of liner.

- Insert wrapped limb into brace frame. Carefully center the uprights over the ankle to ensure a comfortable fit and proper support.




NOTE: Remove plastic covers from uprights and secure hook on uprights to liner.

- Begin fastening the boot straps from the top of foot and working up the leg, using loop lock closures.



- To inflate liner, turn the inflate/deflate dial clockwise. Do not overtighten. Next, pump inflator bulb until desired compression is achieved.
- To deflate liner, turn the inflate/deflate dial counter-clockwise to release air pressure. Turn inflate/deflate dial clockwise when desired compression is achieved.



- Check for adequate circulation. Readjust air amount as needed for a snug and comfortable fit.

WASHING INSTRUCTIONS

Hand wash, lukewarm water, mild detergent, air dry.



Manufactured by DeRoyal
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part#1-1578
©2009 DeRoyal. All Rights Reserved.
Revised 6/09