

# SleepSafe Hi-Lo

## (LOW BED) ASSEMBLY GUIDE



Thank you for purchasing the SleepSafe Safety Bed. With proper care, your bed will provide years of safe use. Carefully unpack all the contents of your bed and verify that all parts are included using the following pages.



**we can help.**  
866-852-2337

**PLEASE CHECK ALL BOXES FOR HARDWARE BEFORE DISCARDING**  
WARRANTY AND USER INFORMATION CAN BE FOUND AT THE END OF THESE INSTRUCTIONS  
*Be sure to run your bed down all the way once a month or multiple resets will be needed.*

### TOOLS NEEDED:



4mm Allen Wrench  
(Provided)



Flat-Head Screwdriver



3/8" Wrench



At least 2 People

# Included Items:



90mm Allen Bolt

**x8**



3/4" Hex Bolt

**x8**



Barrel Nut  
*(Preinstalled on padded beds)*

**x8**



Cardboard Spacer

**x4**



1/4" Allen Wrench

**x1**



4mm Allen Wrench

**x1**



T15 Torx Wrench

**x1**



T10 Torx Wrench

**x1**

# Included Items: (continued)



Casters

**x4**



Hi-Lo Electric Frame

**x1**



Mattress

**x1**



Headboard

**x1**

Footboard

**x1**



Side Rails


**x2**



Safety Rails

**x2**

# Assembly Guide




**1**

## Prepare Side-Rails (insert Barrel Nuts):

Align Barrel Nut Slot Perpendicular to the edge of the Side Rail.

**2**

## Set Frame on Cardboard Spacers:



**3**

## Attach Side Rails to Frame:





# Assembly Guide (continued)



**4**

**Attach Headboard and Footboard to Side Rails:**



**5**

**Install Casters:**




\* Once the caster is inserted, turn the tightening nut as shown above to compress the rubber sleeve on the stem inside the wood leg, thus preventing the caster from falling out.

# Assembly Guide (continued)



## 6 Remove Spacers & Install Mattress:



## 7 Install Safety Rails:



\* Be sure to run your Hi-Lo Bed all the way down once a month or multiple resets will be needed!

# Safety Rail Adjustment

## (UP AND DOWN)



\*Each Safety Rail is pre-adjusted at our factory, but it may become necessary to adjust the Safety Rail height over time.

### TOOLS NEEDED:



T10 Torx Wrench  
(Provided)



T15 Torx Wrench  
(Provided)



Rubber Mallet  
(or Hammer)

## 1 Determine if your Safety Rails need adjustment:

Look at each brass locking-pin when the Safety Rail is in the locked position as indicated by "A" in Fig. 1. The pin should be centered in the hole of the plastic strike. If it is not, note which way the rail needs to move (up or down).




Fig. 1

## 2 Remove Plastic Strikes:

Using your Hammer or Rubber Mallet, gently tap the plastic strike out of place and remove it from the leg.




Fig. 2

## 3 Adjust Accordingly:

Insert the T10 Torx Wrench as shown in Fig. 2. Adjust the screw up or down depending on which way you noted in Step 1.

*(Clockwise = Up / Counterclockwise = Down)*

## 4 Reinsert Plastic Strike:

Place the Plastic Strike back into position and align by gently tapping it with your Rubber Mallet.



# Safety Rail Adjustment

(SIDE TO SIDE)



\*Each Safety Rail is pre-adjusted at our factory, but it may become necessary to adjust the Safety Rail height over time.

## 1 Determine if your Safety Rails need adjustment:


With the safety rail in the locked position, check the spacing between the end of the safety rail and the headboard and footboard legs.

***The spacing should be even!***




## 2 Remove Safety Rail and Adjust Accordingly:

If spacing is not even, or there is too much side-to-side play in the rail, remove safety rail and turn the black hex bolt using the  $\frac{1}{4}$ " Allen Wrench provided.



*B = Clockwise = Increase Play*  
*A = Counterclockwise = Decrease Play*





# The Hi-Lo Remote



**Hi Lo Bed Surface**  
(Up/Down)

**Knee & Foot**  
(Up/Down)

## REMOTE LOCK OUT PROCEDURE:

PRESS and HOLD BOTH DOWN BUTTONS on HEAD and FOOT articulation functions for about 5 seconds. The light on the remote will flash when it is locked.

**Head**  
(Up/Down)

## RESET PROCEDURE:

1. Press and hold the DOWN button on the HiLo function, after the bed is down completely, continue to hold for 5-10 seconds. You will hear the control box click.
2. Your bed is now reset and all functions operate normally.

# Warranty Information



## **Full One-Year Warranty:**

The SleepSafe™ Bed is warranted against defects in workmanship or materials for a period of one year from the date of purchase. SleepSafe Beds will repair any defective part at no cost to the purchaser.



## **Limited Five-Year Warranty:**

During the second through fifth years from the date of purchase, SleepSafe Beds will replace any part found to be defective. Purchaser shall pay any service and shipping costs related to the replacement of the defective part.



## **Terms and Conditions:**

This warranty applies to normal use and does not cover any damage caused by normal wear, abuse, misuse, mishandling or modification of the product.

\*Safety Rails are intended to hinder or discourage the possibility of a roll-out. This product does not take the place of proper client supervision and monitoring. Any modification to the product or its components will void the warranty and render the product incapable of fulfilling its intended purpose.





